

N-Light™ *Listeria monocytogenes*

Product description



Qualitative test for rapid detection of the foodborne bacterial pathogen *Listeria monocytogenes*. The test method is suitable for use in food processing areas and equipment as part of an environmental monitoring program. To be used with the NEMIS Dry block heater and NEMIS Luminometer.

Components

- N-Light™ *L. monocytogenes* Enrichment Broth – 2 ml per test
- N-Light™ *L. monocytogenes* Reagent Tablet – 1 pc. / 13 mg per test



Storage conditions: 2-8 °C / 36-46 °F

Transport conditions: ambient [short duration only]

HS tariff number 38221900
Country of origin CH

Technical specifications

NEMIS tests each production batch against the specifications listed below.

General characteristics	Specification
Broth general appearance	Yellowish, clear waterlike solution
Broth visible turbidity	No particles
Broth optical density (A _{600nm})	< 0.10
Broth Volume	1.80 – 2.20 ml
Broth pH value	6.70 – 7.20
Functionality test (Background)	< 10'000 RLU (NEMIS Luminometer)
Quantity of AquaSpark®	Internal specification

Microbial performance	Specification
Sterility control	No growth, 7 days @ 37 °C
<i>L. monocytogenes</i> ATCC 13932	Positive result
<i>L. monocytogenes</i> ATCC 19111	Positive result
<i>L. ivanovii</i> WSLC 3009	Negative result
<i>E. faecalis</i> ATCC 29212	Negative result
<i>S. aureus</i> ATCC 25923	Negative result

Certificates of analysis are available on request.

N-Light™ *Listeria monocytogenes*

Packaging specifications

General characteristics	Primary packaging [test tube]	Secondary packaging [product box]	Tertiary packaging [distributor box]
Amount	1 test	50 tests [1 x 50]	300 tests [6 x 50]
Bar code / GTIN	See below	7649994979201	7649994979218
Length	1.9 cm	26.5 cm	36.3 cm
Width	1.9 cm	11.8 cm	27.4 cm
Height	11.1 cm	11.5 cm	24.4 cm
Gross weight	0.0095 kg	0.600 kg	3.950 kg
Total plastic amount	0.007 kg	0.350 kg	2.100 kg
Non-reusable plastic amount	0.007 kg	0.350 kg	2.100 kg

Lot and serial number of the test

Each test has a unique number printed on the tube label. The human readable format is LLLLLLL-XXXXX. Additionally, this number is encoded without the hyphen in a linear barcode of type Code 128:

- LLLLLLL: 7-digit lot number
- XXXXX: 5-digit consecutive number

Additional information

Description	Document ID	Availability
Sales brochure	n/a	www.nemistech.com
IFU	Q10041	www.nemistech.com
Material Safety Data Sheet [MSDS]	Q00010	www.nemistech.com
AOAC certificate	Q00022	www.nemistech.com
Shipping and storage conditions for N-Light™ tests	Q00077	www.nemistech.com
FAQ	n/a	www.nemistech.com
Halal statement	Q00198	on request, techsupport@nemistech.com
Phage cleaning validation report	Q00025	on request, techsupport@nemistech.com
Certificates of analysis [CoA's]	Q00048	on request, techsupport@nemistech.com