

# **CERTIFICATION**

# AOAC Research Institute Performance Tested Methods<sup>SM</sup>

Certificate No.

122002

The AOAC Research Institute hereby certifies the method known as:

# N-Light<sup>™</sup> *Listeria monocytogenes*

manufactured by
NEMIS Technologies AG
Riedhofstrasse 11
CH - 8804 Au, ZH
Switzerland

This method has been evaluated in the AOAC Research Institute *Performance Tested Methods*<sup>SM</sup> Program and found to perform as stated in the applicability of the method. This certificate indicates an AOAC Research Institute Certification Mark License Agreement has been executed which authorizes the manufacturer to display the AOAC Research Institute *Performance Tested Methods* SM certification mark on the above-mentioned method for the period below. Renewal may be granted by the Expiration Date under the rules stated in the licensing agreement.

Bradley A. Stawick, Senior Director Signature for AOAC Research Institute

Issue Date
Expiration Date

February 16, 2024 December 31, 2024

#### **AUTHORS**

ORIGINAL MODIFICATION: Delphine Larose, Nicolas Desroches, Lukas Reinau, Lars Fieseler, Mario Hupfeld

MODIFICATION JANUARY 2023: Nicolas Desroche<sup>1</sup>, Déborah De Oliveira<sup>1</sup>,

Lukasz Richter<sup>2</sup>, and Mario Hupfeld<sup>2</sup>

MODIFICATION FEBRUARY 2024: Wesley Thompson<sup>3</sup>, M. Joseph Benzinger, Jr.<sup>3</sup>, Benjamin Bastin<sup>3</sup>, Erin S. Crowley<sup>3</sup>, Nicolas Desroche<sup>1</sup>, Julian Ihssen<sup>2</sup>, and Mario Hupfeld<sup>2</sup>

- <sup>1</sup>Nexidia SAS, 15 Rue de Mayence, 21000 Dijon, FRANCE
- <sup>2</sup> Nemis Technologies AG, Riedhofstrasse 11, 8804 Au (ZH) SWITXERLAND
- <sup>3</sup> Q Laboratories, Cincinnati, OH USA

SUBMITTING COMPANY NEMIS Technologies AG Riedhofstrasse 11 CH - 8804 Au, ZH Switzerland

## METHOD NAME

N-Light<sup>™</sup> *Listeria monocytogenes* 

#### CATALOG NUMBERS

00009 (50 tests), 00010 (Distributor Box)

Retired: 00005 (48 tests)

#### INDEPENDENT LABORATORY

Q Laboratories 1930 Radcliff Drive Cincinnati, OH 45204 USA

# CONTRACT LABORATORY MODIFICATION JANUARY 2023

Nexidia Laboratories 15 Rue de Mayence 21000 Dijon, France

## APPLICABILITY OF METHOD

Analytes - Listeria monocytogenes.

Matrixes – Environmental Surface Swabs (1" x 1"): Stainless steel [AISI 304 (1.4301), grade 2b finish], plastic (polystyrene), and Ceramic (glazed earthen)

Performance claims – No significant statistical difference was detected between the N-Light *Listeria monocytogenes* assay, using the NEMIS Technologies BTL1 luminometer, and the ISO 11290-1:2017: Microbiology of the Food Chain – Horizontal method for the detection and enumeration of Listeria monocytogenes and of Listeria spp. – Part 1 (2) for detection of *L. monocytogenes* on stainless steel environmental surfaces.

#### **ORIGINAL CERTIFICATION DATE**

December 07, 2020

#### CERTIFICATION RENEWAL RECORD

Renewed annually through December 2024.

## METHOD MODIFICATION RECORD

1. December 2021 Level 1

February 2024 Level 3

2. January 2023 Level 2

#### SUMMARY OF MODIFICATION

- 1. Editorial and formatting changes updated catalog numbers.
- 2. Modification changes include:
  - Changes to the AquaSpark® Tablet excipients for lower background and improved signal,
  - Option to use swabs with a breaking point that can be left in the enrichment tube during incubation,
  - Reduction in reading time of luminescence signal after tablet release reduced to 5 minutes.
- 3. Modification changes include:
  - a. New proprietary premoistening buffer,
  - Enrichment broth was optimized in terms of phage cocktail and nutrient composition.

Under this AOAC Performance Tested Methods  $^{\rm SM}$  License Number, 122002 this method is distributed by:

NONE

Under this AOAC *Performance Tested Methods*<sup>SM</sup> License Number, 122002 this method is distributed as:

NON

#### PRINCIPLE OF THE METHOD (1)

NEMIS Technologies N-light™ *L. monocytogenes* (L. mono) assay utilizes synthetic molecules (AquaSpark™) that are designed to detect live bacteria through chemiluminescence. The mechanism of detection is the hydrolysis of the proprietary AquaSpark molecule by *L. monocytogenes* specific enzymatic cleavage (PI-PLC). This leads to the release of a dioxetane derivate which in turn leads to the emission of green light. The presence of *L. monocytogenes* is determined by the relative light unit (RLU) reading for the sample. If the sample eclipses the RLU threshold, which is set to 20,000 RLU, then the sample is deemed positive for the presence of *L. monocytogenes*. Unlike molecular methods which require a DNA purification or lysis process, the chemiluminescence detection has a simple, streamlined sampling protocol. After swabbing and incubating the enrichment for 24 h (dry bath incubator), the end user only has to dispense the AquaSpark tablet, wait 10 minutes and insert the tube containing the enrichment into the NEMIS Luminometer for detection in 10 seconds. NEMIS uses its own proprietary enrichment broth contining phage selectives which specifically reduces the background flora in particular *Enterococci, Staphylococci* and non-*monocytogenes Listeria* spp.

#### **DISCUSSION OF THE ORIGINAL VALIDATION STUDY (1)**

The N-Light L. mono kit was able to detect all 52 of the *L. monocytogenes* strains, representing 12 different serotypes, tested during the inclusivity study. Moreover, the N-light L. mono kit did not detect any of the 30 non-*Listeria monocytogenes* in the exclusivity strain set, which included other *Listeria* spp. The specificity of the kit was therefore validated according to the exclusivity study.

Concerning the matrix study, on stainless steel without competitive flora, the N-light L. mono method able to detect significantly more positive sample than the reference method for the low inoculation, for stainless steel and ceramics (see Table 5). For plastics on the other hand, significantly more positives were detected with the reference method. When stainless steel surface was co-inoculated with L. monocytogenes and E. faecalis, significantly more positives were detected using the N-light method compared to the reference method. It must be noted that environmental swabbing on surfaces is subject to high variation in between experimental days, therefore results need to be treated with the respective caution.

Table 1:	Inclusivity Panel Results (1)					
#	Organism	Serotype	Source <sup>a</sup>	Strain	Origin	Results
1	Listeria monocytogenes	1/2a	ATCC	15313	Rabbit	Positive
2	Listeria monocytogenes	1/2a	Nexidia	NEX-636	Cheese	Positive
3	Listeria monocytogenes	1/2a	Nexidia	NEX-718	Food isolate	Positive
4	Listeria monocytogenes	1/2a	Nexidia	NEX-726	Food isolate	Positive
5	Listeria monocytogenes	1/2a	Nexidia	NEX-772	Cheese	Positive
6	Listeria monocytogenes	1/2a	Nexidia	NEX-773	Cheese	Positive
7	Listeria monocytogenes	1/2a	Nexidia	NEX-789	Cheese	Positive
8	Listeria monocytogenes	1/2a	Nexidia	NEX-865	Food isolate	Positive
9	Listeria monocytogenes	1/2a	Nexidia	NEX-866	Food isolate	Positive
10	Listeria monocytogenes	1/2a	Nexidia	NEX-939	Food isolate	Positive
11	Listeria monocytogenes	1/2a	Nexidia	NEX-952	Food isolate	Positive
12	Listeria monocytogenes	1/2a	Nexidia	NEX-1492	Food isolate	Positive
13	Listeria monocytogenes	1/2a	Nexidia	NEX-1494	Food isolate	Positive
14	Listeria monocytogenes	1/2a	Nexidia	NEX-1495	Food isolate	Positive
15	Listeria monocytogenes	1/2a	Nexidia	NEX-1497	Food isolate	Positive
16	Listeria monocytogenes	1/2a	Nexidia	NEX-1498	Food isolate	Positive
17	Listeria monocytogenes	1/2a	Nexidia	NEX-1506	Food isolate	Positive
18	Listeria monocytogenes	1/2b	CIP	78.32 (= 105449)	Chinchilla	Positive
19	Listeria monocytogenes	1/2b	CIP	78.33 (= 105448)	Cerebrospinal fluid	Positive
20	Listeria monocytogenes	1/2b	Nexidia	NEX-626	Food isolate	Positive
21	Listeria monocytogenes	1/2b	Nexidia	NEX-648	Food isolate	Positive
22	Listeria monocytogenes	1/2b	Nexidia	NEX-720	Cheese	Positive
23	Listeria monocytogenes	1/2b	Nexidia	NEX-841	Food isolate	Positive
24	Listeria monocytogenes	1/2b	Nexidia	NEX-851	Food isolate	Positive
25	Listeria monocytogenes	1/2b	Nexidia	NEX-1493	Food isolate	Positive
26	Listeria monocytogenes	1/2b	Nexidia	NEX-870	Food isolate	Positive
27	Listeria monocytogenes	1/2b	Nexidia	BAA-751	-	Positive
28	Listeria monocytogenes	1/2c	Nexidia	NEX-777	Food isolate	Positive
29	Listeria monocytogenes	1/2c	Nexidia	NEX-869	Food isolate	Positive
30	Listeria monocytogenes	1/2c	Nexidia	NEX-868	Food isolate	Positive
31	Listeria monocytogenes	1/2c	ATCC	7644	Human	Positive
32	Listeria monocytogenes	3a	CIP	78.34	Human	Positive
33	Listeria monocytogenes	3a	Nexidia	NEX-627	Food isolate	Positive
34	Listeria monocytogenes	3a	Nexidia	NEX-1496	Food isolate	Positive
35	Listeria monocytogenes	3a	Nexidia	NEX-1505	Food isolate	Positive
36	Listeria monocytogenes	3b	CIP	78.35	Spinal fluid	Positive
37	Listeria monocytogenes	3b	CRBIP	13.109	Environment	Positive
38	Listeria monocytogenes	3c	CIP	78.36	-	Positive
39	Listeria monocytogenes	3c	CRBIP	13.36	Meat product	Positive
40	Listeria monocytogenes	4a	NCTC	5214	Brain sheep	Positive
41	Listeria monocytogenes	4b	CIP	78.39	Chicken	Positive
42	Listeria monocytogenes	4b	CRBIP	13.107	Environment	Positive
43	Listeria monocytogenes	4b	Nexidia	NEX-642	Cheese	Positive
44	Listeria monocytogenes	4b	Nexidia	NEX-754	Food isolate	Positive
45	Listeria monocytogenes	4b	Nexidia	NEX-929	Food isolate	Positive
46	Listeria monocytogenes	4b	Nexidia	NEX-1504	Food isolate	Positive
47	Listeria monocytogenes	4c	ATCC	19116	Chicken	Positive
48	Listeria monocytogenes	4d	ATCC	19118	Sheep	Positive
49	Listeria monocytogenes	4d	NCTC	10888	Sheep	Positive
50	Listeria monocytogenes	4e	CIP	78.41	Chicken	Positive
51	Listeria monocytogenes	7	Nexidia	NEX-843	Food isolate	Positive
52	Listeria monocytogenes	7	CIP	78.43	-	Positive

<sup>a</sup>ATCC: American Type Culture Collection, Manassas, VA, USA. CIP: Collection de l'Institut Pasteur, Paris, France. CRBIP: Biological Resource Center of institute Pasteur, Paris, France. NCTC: National Collection of Type Cultures, Salisbury, UK. NEXIDIA: Microbial Strain Collection, Dijon, France.

Table 2:	Exclusivity Panel Results (1)					
#	Organism	Serotype	Sourcea	Strain	Origin	Results
1	Listeria innocua	-	ATCC	33090	Cow brain	Negative
2	Listeria innocua	-	Nexidia	NEX-1849	Milk	Negative
3	Listeria ivanovii	-	ATCC	BAA-139	Washing water	Negative
4	Listeria ivanovii	-	Nexidia	NEX-923	Duck breast	Negative
5	Listeria grayi	-	CIP	103321	Food	Negative
6	Listeria seeligeri	-	CIP	100100	Soil	Negative
7	Listeria seeligeri	-	Nexidia	NEX-1848	Milk	Negative
8	Listeria welshimerii	-	CIP	81.49	Decaying vegetation	Negative
9	Lactococcus lactis	-	ATCC	11454	-	Negative
10	Lactococcus lactis	-	Nexidia	NEX-363	-	Negative
11	Streptococcus thermophilus	-	Nexidia	NEX-371	Food isolate	Negative
12	Lactobacillus plantarum	-	ATCC	14917	Pickled cabbage	Negative
13	Lactobacillus plantarum	-	Nexidia	NEX-EC-EnrE8	Surface	Negative
14	Lactobacillus buchneri	-	Nexidia	NEX-EC-E8	Surface	Negative
15	Klebsiella pneumoniae	-	ATCC	4352	Dairy products: cow's milk	Negative
16	Enterobacter amnigenus	-	Nexidia	NEX-764	Cocoa	Negative
17	Cronobacter sakazakii	-	ATCC	12868	-	Negative
18	Escherichia coli	-	ATCC	8739	Faeces	Negative
19	Citrobacter braakii	-	ATCC	51113	Snake	Negative
20	Pantoea agglomerans	-	CIP	82.100	Corn crops	Negative
21	Salmonella Typhimurium	-	ATCC	14028	Chicken	Negative
22	Pseudomonas aeruginosa	-	ATCC	15442	Animal room water bottle	Negative
23	Enterococcus avium	-	ATCC	14025	-	Negative
24	Enterococcus faecalis	-	ATCC	51299	Peritoneal fluid	Negative
25	Enterococcus hirae	-	ATCC	10541	-	Negative
26	Staphylococcus aureus	-	Nexidia	NEX-1093	Raw milk cheese	Negative
27	Staphylococcus aureus	-	Nexidia	NEX-1069	Mayonnaise	Negative
28	Bacillus cereus	-	CIP	105151	-	Negative
29	Bacillus subtilis	-	ATCC	6051	-	Negative
30	Clostridium perfringens	-	ATCC	13124	-	Negative

<sup>a</sup>ATCC: American Type Culture Collection, Manassas, VA, USA. CIP: Collection de l'Institut Pasteur, Paris, France. CRBIP: Biological Resource Center of institute Pasteur, Paris, France. NCTC: National Collection of Type Cultures, Salisbury, UK. NEXIDIA: Microbial Strain Collection, Dijon, France.

Table 4: NE	MIS Technologies N-ligl	nt <i>L. mono</i> Assa	ay: Pro	esum	ptive vs. Co	onfirmed (1)					
Matrix	Strain	CFU/Test	N <sup>b</sup>			te Method Imptive	Candio	date Metho	d Confirmed	$dPOD_{CP}^f$	95% CI <sup>g</sup>
		Area		Х <sup>с</sup>	$POD_{CP^d}$	95% CI	Х	POD <sub>cc</sub> e	95% CI		
Cha: alass		0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.43,0.43)
Stainless steel	Listeria monocytogenes ATCCh BAA-751 (1/2b)	72	20	19	0.95	(0.76, 0.99)	19	0.95	(0.76, 0.99)	0.00	(-0.13,0.13)
Steel	ATCC BAA-731 (1/20)	2 200	5	5	1.00	(0.57,1.00)	5	1.00	(0.57,1.00)	0.00	(-0.43,0.43)
	Listeria monocytogenes	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.47,0.47)
Stainless	ATCC BAA-751	73 & 810	20	8	0.40	(0.22, 0.61)	8	0.40	(0.22, 0.61)	0.00	(-0.13,0.13)
steel <sup>i</sup>	&Enterococcus faecalis ATCC 29212	220 & 2400	5	5	1.00	(0.57,1.00)	5	1.00	(0.57,1.00)	0.00	(-0.47,0.47)
	Listeria	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.43,0.43)
Plastic	monocytogenes	115	20	11	0.55	(0.34,0.74)	11	0.55	(0.34, 0.74)	0.00	(-0.13,0.13)
	ATCC 7644 (1/2c)	10 000	5	4	0.8	(0.38,0.96)	4	0.80	(0.38,0.96)	0.00	(-0.45,0.45)
	Listeria	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.43,0.43)
Ceramic	monocytogenes	210	20	16	0.8	(0.58,0.92)	16	0.8	(0.58,0.92)	0.00	(-0.13,0.13)
ceranne	ATCC 19118 (4d)	22 000	5	5	1.00	(0.57,1.00)	5	1.00	(0.57,1.00)	0.00	(-0.43,0.43)

<sup>&</sup>lt;sup>a</sup>CFU/Test area determined by plating the inoculum in duplicate (shown as average)

<sup>&</sup>lt;sup>b</sup>N = Number of test potions.

cx = Number of positive test portions.

<sup>&</sup>lt;sup>d</sup>POD<sub>CP</sub> = Candidate method presumptive positive outcomes divided by the total number of trials.

<sup>&</sup>lt;sup>e</sup>POD<sub>CC</sub> = Candidate method confirmed positive outcomes divided by the total number of trials.

<sup>&</sup>lt;sup>f</sup>dPOD<sub>CP</sub> = Difference between the candidate method presumptive result and candidate method confirmed result POD values.

<sup>895%</sup> CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level.

<sup>&</sup>lt;sup>h</sup> ATCC = American Type Culture Collection, Manassas, VA, USA

<sup>&</sup>lt;sup>1</sup> Performed by independent AOAC certified laboratory Q Laboratories, Cincinnati, OH, USA

Matrix	Strain	CFU/Test	$N^b$	Can	didate me (confir	thod results med)	ISO 11	1290-1 resul	ts (confirmed)	$dPOD_{\mathcal{C}^f}$	95% CI <sup>g</sup>
		Area		X <sup>c</sup>	$POD_{C^d}$	95% CI	Х	$POD_R^e$	95% CI		
	Listeria	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.47,0.47)
Stainless	monocytogenes	72	20	19	0.95	(0.76, 0.99)	15	0.75	(0.53, 0.89)	0.20	(-0.03, 0.42)
steel	ATCC <sup>h</sup> BAA-751 (1/2b)	2 200	5	5	1.00	(0.57,1.00)	5	1.00	(0.57,1.00)	0.00	(-0.47,0.47)
	Listeria	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.43, 0.43)
Stainless steel <sup>i</sup>	monocytogenes	73 & 810	20	8	0.4	(0.22, 0.61)	6	0.30	(0.15, 0.52)	0.10	(-0.18, 0.36)
	ATCC BAA-751 &Enterococcus faecalis ATCC 29212	220 & 2400	5	5	1.00	(0.57,1.00)	5	1.00	(0.57,1.00)	0.00	(-0.43, 0.43)
	Listeria	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.47,0.47)
Plastic	monocytogenes	115	20	11	0.55	(0.34, 0.74)	14	0.70	(0.48,0.85)	-0.15	(-0.41,0.14)
	ATCC 7644 (1/2c)	10 000	5	4	0.80	(0.38,0.96)	5	1.00	(0.57,1.00)	-0.20	(-0.62,0.28)
	Listeria	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.47,0.47)
Ceramic	monocytogenes	210	20	16	0.80	(0.58, 0.92)	12	0.60	(0.39,0.78)	0.20	(-0.08,0.44)
	ATCC 19118 (4d)	22 000	5	5	1.00	(0.57, 1.00)	4	0.8	(0.38, 0.96)	0.20	(-0.28, 0.62)

<sup>&</sup>lt;sup>a</sup>CFU/Test area determined by plating the inoculum in triplicate (shown as average)

#### **DISCUSSION OF MODIFICATION APPROVED JANUARY 2023 (3)**

Since the original validation (Certificate No. 122002 – December 11<sup>th</sup> 2020), the N-Light *Listeria monocytogenes* method for the detection of *L. monocytogenes* was modified to improve the detection on surfaces. The *N-light L. monocytogenes* kit was able to detect all the *L. monocytogenes* strains, covering serotypes 1/2a, 1/2b, 1/2c, 3a, 3b, 3c, 4a, 4b, 4c, 4d, 4e and 7, during the inclusivity study. Moreover, the N-Light *L. monocytogenes* kit did not detect any of the exclusivity strains, which included non-*monocytogenes Listeria* species in the exclusivity strain set.

In the matrix study on stainless steel with competitor micro-organism (*E. faecalis*), the N-Light *L. monocytogenes* method detected significantly more positive test portions in fractional results compared to the ISO 11290-1.

During the robustness study, changes to incubation time, time of incubation between tablet release and measurement, and method for swab agitation did not show any statistical differences in terms of assay performance. However, the time of incubation between tablet release and signal measurement may affect the signal values and will be underlined in the method kit instructions.

The accelerated stability study demonstrated that shelf-life exposure does not change the POD in a statistically significant manner for the detection of positive signals from *L. monocytogenes*. However, after 24 days of storage at 37°C, a significant increase of the signal was observed for the competitive micro-organism. Therefore, a shelf life of longer than 6 months will need to be proven by real time measurements at the recommended storage temperature (2-8°C).

Table 1	L: Inclusivity Panel Results (3)					
No.	Organism	Serotype	Source <sup>a</sup>	Strain	Origin	Results
1	Listeria monocytogenes	1/2a	ATCC	15313	Rabbit	Positive
2	Listeria monocytogenes	1/2a	Nexidia	NEX-636	Cheese	Positive
3	Listeria monocytogenes	1/2a	Nexidia	NEX-772	Cheese	Positive
4	Listeria monocytogenes	1/2a	Nexidia	NEX-773	Cheese	Positive
5	Listeria monocytogenes	1/2a	Nexidia	NEX-789	Cheese	Positive
6	Listeria monocytogenes	1/2a	Nexidia	NEX-866	Food isolate	Positive
7	Listeria monocytogenes	1/2a	Nexidia	NEX-952	Food isolate	Positive
8	Listeria monocytogenes	1/2a	Nexidia	NEX-1495	Food isolate	Positive
9	Listeria monocytogenes	1/2b	CIP	78.32 (= 105449)	Chinchilla	Positive
10	Listeria monocytogenes	1/2b	CIP	78.33 (= 105448)	Cerebrospinal fluid	Positive
11	Listeria monocytogenes	1/2b	Nexidia	NEX-648	Food isolate	Positive
12	Listeria monocytogenes	1/2b	Nexidia	NEX-720	Cheese	Positive
13	Listeria monocytogenes	1/2b	Nexidia	NEX-1493	Food isolate	Positive
14	Listeria monocytogenes	1/2c	Nexidia	NEX-869	Food isolate	Positive
15	Listeria monocytogenes	1/2c	ATCC	7644	Human	Positive
16	Listeria monocytogenes	3a	CIP	78.34	Human	Positive
17	Listeria monocytogenes	3b	CIP	78.35	Spinal fluid	Positive
18	Listeria monocytogenes	3b	CRBIP	13.109	Environment	Positive
19	Listeria monocytogenes	3c	CRBIP	13.36	Meat product	Positive
20	Listeria monocytogenes	4a	NCTC	5214	Brain sheep	Positive
21	Listeria monocytogenes	4b	CIP	78.39	Chicken	Positive

<sup>&</sup>lt;sup>b</sup>N = Number of test potions.

 $<sup>^{</sup>c}x$  = Number of positive test portions.

<sup>&</sup>lt;sup>d</sup>POD<sub>C</sub> = Candidate method presumptive positive outcomes confirmed positive divided by the total number of trials.

<sup>&</sup>lt;sup>e</sup>POD<sub>R</sub> = Reference method confirmed positive outcomes divided by the total number of trials.

fdPODc = Difference between the candidate method and reference method POD values.

<sup>995%</sup> CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level.

h ATCC = American Type Culture Collection, Manassas, VA, USA

<sup>&</sup>lt;sup>1</sup> Performed by independent AOAC certified laboratory Q Laboratories, Cincinnati, OH, USA

### NEMIS Technologies AG N-Light™ Listeria monocytogenes, AOAC Performance Tested Methods<sup>SM</sup> certification number 122002

22	Listeria monocytogenes	4b	CRBIP	13.107	Environment	Positive
23	Listeria monocytogenes	4b	Nexidia	NEX-642	Cheese	Positive
24	Listeria monocytogenes	4c	ATCC	19116	Chicken	Positive
25	Listeria monocytogenes	4d	ATCC	19118	Sheep	Positive
26	Listeria monocytogenes	4e	CIP	78.41	Chicken	Positive
27	Listeria monocytogenes	7	Nexidia	NEX-843	Food isolate	Positive

<sup>&</sup>lt;sup>a</sup>American Type Culture Collection (ATCC), Manassas, VA, USA; Collection de l'Institut Pasteur (CIP), Paris, France; National Collection of Type Cultures (NCTC), Salisbury, United Kingdom; Nexidia Microbial Strain Collection (NEXIDIA), Dijon, France.

Table 2	: Exclusivity Panel Results (3)				
No.	Organism	Sourcea	Strain	Origin	Results
1	Listeria innocua	ATCC	33090	Cow brain	Negative
2	Listeria innocua	Nexidia	NEX-1849	Milk	Negative
3	Listeria ivanovii	ATCC	BAA-139	Washing water	Negative
4	Listeria ivanovii	Nexidia	NEX-923	Duck breast	Negative
5	Listeria grayi	CIP	103321	Food	Negative
6	Listeria seeligeri	CIP	100100	Soil	Negative
7	Listeria seeligeri	Nexidia	NEX-1848	Milk	Negative
8	Listeria welshimerii	CIP	81.49	Decaying vegetation	Negative
9	Klebsiella pneumoniae	ATCC	4352	Dairy products: cow's milk	Negative
10	Enterobacter amnigenus	Nexidia	NEX-764	Cocoa	Negative
11	Escherichia coli	ATCC	8739	Feces	Negative
12	Citrobacter braakii	ATCC	51113	Snake	Negative
13	Pantoea agglomerans	CIP	82.100	Corn crops	Negative
14	Salmonella Typhimurium	ATCC	14028	Chicken	Negative
15	Enterococcus faecalis	ATCC	51299	Peritoneal fluid	Negative

<sup>&</sup>lt;sup>a</sup>American Type Culture Collection (ATCC), Manassas, VA, USA; Nexidia Microbial Strain Collection (NEXIDIA), Dijon, France; Collection de l'Institut Pasteur (CIP), Paris, France.

Table 3. N-L	Table 3. N-Light <i>Listeria monocytogenes</i> : Presumptive vs. Confirmed (with enrichments in Fraser-Demi and Fraser) (3)										
		cfu/Test		Candidate method presumptive			Candidate method confirmed				
Matrix	Strain	area	$N^b$	Xc	$POD_{CP^d}$	95% CI	х	POD <sub>cc</sub> e	95% CI	$dPOD_{CP}^f$	95% CI <sup>g</sup>
	L. monocytogenes	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.47,0.47)
Stainless steel	ATCC BAA-751 & E. faecalis	17 & 637	20	15	0.75	(0.53,0.89)	16	0.80	(0.58,0.92)	-0.05	(-0.21,0.11)
	ATCC 29212	140 & 8 630	5	5	1.00	(0.57,1.00)	5	1.00	(0.57,1.00)	0.00	(-0.47,0.47)

<sup>&</sup>lt;sup>a</sup>cfu/Test area determined by plating the inoculum in duplicate

<sup>&</sup>lt;sup>h</sup> ATCC = American Type Culture Collection, Manassas, VA, USA.

Table 4. N-I	Table 4. N-Light Listeria monocytogenes: Presumptive vs. Alternative Confirmed (Direct Streaking onto Agar Plates) (3)										
		cfu/Test				e method nptive	Can	didate met	hod confirmed		
Matrix	Strain	area	$N^b$	xc	$POD_{CP^d}$	95% CI	х	PODcce	95% CI	$dPOD_{CP}^{f}$	95% CI <sup>g</sup>
	L. monocytogenes	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.47,0.47)
Stainless steel	ATCC BAA-751 & E. faecalis	17 & 637	20	15	0.75	(0.53,0.89)	16	0.80	(0.58,0.92)	-0.05	(-0.21,0.11)
	ATCC 29212	140 & 8 630	5	5	1.00	(0.57,1.00)	5	1.00	(0.57,1.00)	0.00	(-0.47,0.47)

<sup>&</sup>lt;sup>a</sup>cfu/Test area determined by plating the inoculum in triplicate.

<sup>&</sup>lt;sup>b</sup>N = Number of test potions.

cx = Number of positive test portions.

<sup>&</sup>lt;sup>d</sup>POD<sub>CP</sub> = Candidate method presumptive positive outcomes divided by the total number of trials.

<sup>&</sup>lt;sup>e</sup>POD<sub>CC</sub> = Candidate method confirmed positive outcomes divided by the total number of trials.

fdPOD<sub>CP</sub> = Difference between the candidate method presumptive result and candidate method confirmed result POD values.

<sup>\$95%</sup> CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level.

<sup>&</sup>lt;sup>b</sup>N = Number of test potions.

 $<sup>^{</sup>c}x$  = Number of positive test portions.

<sup>&</sup>lt;sup>d</sup>POD<sub>CP</sub> = Candidate method presumptive positive outcomes divided by the total number of trials.

<sup>&</sup>lt;sup>e</sup>POD<sub>CC</sub> = Candidate method confirmed positive outcomes divided by the total number of trials.

fdPOD<sub>CP</sub> = Difference between the candidate method presumptive result and candidate method confirmed result POD values.

<sup>895%</sup> CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level.

Table 5. Method Comparison Results: N-Light Listeria monocytogenes (Traditional Confirmation) vs. ISO 11290-1:2017 (3)

		cfu/Test		Can	didate Met	hod Confirmed		ISO 112	90-1:2017	=	
Matrix	Strain	Area	N <sup>b</sup>	Χ <sup>c</sup>	$POD_{CP^d}$	95% CI	х	PODcce	95% CI	$dPOD_{CP}^f$	95% CI <sup>g</sup>
	L. monocytogenes	0	5	0	0.00	(0.00,0.43)	0	0.00	(0.00,0.43)	0.00	(-0.43,0.43)
Stainless steel	ATCC <sup>h</sup> BAA-751 & <i>E. faecalis</i>	17 & 637	20	15	0.75	(0.53,0.89)	6	0.30	(0.15,0.52)	0.45	(0.14,0.66)
	ATCC 29212	140 & 8 630	5	5	1.00	(0.57,1.00)	5	1.00	(0.57,1.00)	0.00	(-0.43,0.43)

<sup>&</sup>lt;sup>a</sup>cfu/Test area determined by plating the inoculum in triplicate.

#### **DISCUSSION OF MODIFICATION APPROVED FEBRUARY 2024 (4)**

An inclusivity/exclusivity study showed that all major serotypes of *L. monocytogenes* are detected with the method while non-*L. monocytogenes Listeria* species and numerous other bacterial species commonly occurring in food factory environments yield negative test results.

The NEMIS Technologies N-Light *Listeria monocytogenes* assay successfully detected *L. monocytogenes* on stainless steel environmental surfaces. When comparing results obtained from the BTL1 luminometer to the confirmation results, no false positives or false negatives were observed. Using POD analysis, no statistically significant differences were observed between the number of positive samples detected by the reference method and the N-Light *Listeria monocytogenes* assay. The method allows the user to swab a surface and obtain next-day results for the presence of *L. monocytogenes*.

The method is quick and easy to perform with the instrument, only taking 10 seconds to provide a result for a sample. This short run time allows an analyst to process multiple samples in a short amount of time. The product manual is well written and simple enough that a technician at any level of training would be able to follow it and achieve accurate results. The BTL1 luminometer is also easy to use and displays the RLU value prominently. The instrument is very small and takes up little bench space, allowing for flexibility in process set up.

Fable 1: Inclusivity Panel Results (4)										
No.	Organism	Serotype	Source	Strain	Origin	Results				
1	Listeria monocytogenes	1/2a	ATCC <sup>a</sup>	15313	Rabbit	Positive				
_	Photo de la colonia de la colonia	1 /2-	N1 - 1 - 11 - h	NEV VA -	Channa	D'1'				

1101	O Barrisin	oc. otype	304.00	Strain	0.16.11	itesaits
1	Listeria monocytogenes	1/2a	ATCC <sup>a</sup>	15313	Rabbit	Positive
2	Listeria monocytogenes	1/2a	Nexidia <sup>b</sup>	NEX-YAe	Cheese	Positive
3	Listeria monocytogenes	1/2a	Nexidia	NEX-718	Food isolate	Positive
4	Listeria monocytogenes	1/2a	Nexidia	NEX-726	Food isolate	Positive
5	Listeria monocytogenes	1/2a	Nexidia	NEX-772	Cheese	Positive
6	Listeria monocytogenes	1/2a	Nexidia	NEX-773	Cheese	Positive
7	Listeria monocytogenes	1/2a	Nexidia	NEX-789	Cheese	Positive
8	Listeria monocytogenes	1/2a	Nexidia	NEX-865	Food isolate	Positive
9	Listeria monocytogenes	1/2a	Nexidia	NEX-866	Food isolate	Positive
10	Listeria monocytogenes	1/2a	Nexidia	NEX-939	Food isolate	Positive
11	Listeria monocytogenes	1/2a	Nexidia	NEX-952	Food isolate	Positive
12	Listeria monocytogenes	1/2a	Nexidia	NEX-1492	Food isolate	Positive
13	Listeria monocytogenes	1/2a	Nexidia	NEX-1494	Food isolate	Positive
14	Listeria monocytogenes	1/2a	Nexidia	NEX-1495	Food isolate	Positive
15	Listeria monocytogenes	1/2a	Nexidia	NEX-1497	Food isolate	Positive
16	Listeria monocytogenes	1/2a	Nexidia	NEX-1498	Food isolate	Positive
17	Listeria monocytogenes	1/2a	Nexidia	NEX-1506	Food isolate	Positive
18	Listeria monocytogenes	1/2b	CIP <sup>c</sup>	78.32 (= 105449)	Chinchilla	Positive
19	Listeria monocytogenes	1/2b	CIP	78.33 (= 105448)	Cerebrospinal fluid	Positive
20	Listeria monocytogenes	1/2b	Nexidia	NEX-626	Food isolate	Positive
21	Listeria monocytogenes	1/2b	Nexidia	NEX-648	Food isolate	Positive
22	Listeria monocytogenes	1/2b	Nexidia	NEX-720	Cheese	Positive
23	Listeria monocytogenes	1/2b	Nexidia	NEX-841	Food isolate	Positive
24	Listeria monocytogenes	1/2b	Nexidia	NEX-851	Food isolate	Positive
25	Listeria monocytogenes	1/2b	Nexidia	NEX-1493	Food isolate	Positive
26	Listeria monocytogenes	1/2b	Nexidia	NEX-870	Food isolate	Positive
27	Listeria monocytogenes	1/2b	Nexidia	BAA-751	Not available	Positive
28	Listeria monocytogenes	1/2c	Nexidia	NEX-777	Food isolate	Positive
29	Listeria monocytogenes	1/2c	Nexidia	NEX-869	Food isolate	Positive
30	Listeria monocytogenes	1/2c	Nexidia	NEX-868	Food isolate	Positive
31	Listeria monocytogenes	1/2c	ATCC	7644	Human	Positive
32	Listeria monocytogenes	3a	CIP	78.34	Human	Positive
33	Listeria monocytogenes	3a	Nexidia	NEX-627	Food isolate	Positive
34	Listeria monocytogenes	3a	Nexidia	NEX-1496	Food isolate	Positive
35	Listeria monocytogenes	3a	Nexidia	NEX-1505	Food isolate	Positive
36	Listeria monocytogenes	3b	CIP	78.35	Spinal fluid	Positive

<sup>&</sup>lt;sup>b</sup>N = Number of test potions.

cx = Number of positive test portions.

<sup>&</sup>lt;sup>d</sup>POD<sub>C</sub> = Candidate method presumptive positive outcomes confirmed positive divided by the total number of trials.

<sup>&</sup>lt;sup>e</sup>POD<sub>R</sub> = Reference method confirmed positive outcomes divided by the total number of trials.

<sup>&</sup>lt;sup>f</sup>dPOD<sub>C</sub> = Difference between the candidate method and reference method POD values.

<sup>895%</sup> CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level.

<sup>&</sup>lt;sup>h</sup>ATCC = American Type Culture Collection, Manassas, VA, USA.

37	Listeria monocytogenes	3b	CRBIPd	13.109	Environment	Positive
38	Listeria monocytogenes	3c	CIP	78.36	Not available	Positive
39	Listeria monocytogenes	3c	CRBIP	13.36	Meat product	Positive
40	Listeria monocytogenes	4a	NCTC	5214	Brain sheep	Positive
41	Listeria monocytogenes	4b	CIP	78.39	Chicken	Positive
42	Listeria monocytogenes	4b	CRBIP	13.107	Environment	Positive
43	Listeria monocytogenes	4b	Nexidia	NEX-642	Cheese	Positive
44	Listeria monocytogenes	4b	Nexidia	NEX-754	Food isolate	Positive
45	Listeria monocytogenes	4b	Nexidia	NEX-929	Food isolate	Positive
46	Listeria monocytogenes	4b	Nexidia	NEX-1504	Food isolate	Positive
47	Listeria monocytogenes	4c	ATCC	19116	Chicken	Positive
48	Listeria monocytogenes	4d	ATCC	19118	Sheep	Positive
49	Listeria monocytogenes	4d	NCTC	10888	Sheep	Positive
50	Listeria monocytogenes	4e	CIP	78.41	Chicken	Positive
51	Listeria monocytogenes	7	Nexidia	NEX-843	Food isolate	Positive
52	Listeria monocytogenes	7	CIP	78.43	Not available	Positive

<sup>&</sup>lt;sup>a</sup> ATCC: American Type Culture Collection, Manassas, VA, USA.

<sup>&</sup>lt;sup>e</sup> Nexidia: Microbial Strain Collection Nexidia SAS, Dijon, France.

Table 2:	Exclusivity Panel Results (4)					
No.	Organism	Source	Strain	Origin	Results	N-Light broth
1	Listeria innocua	ATCCa	33090	Cow brain	Negative	Not tested
2	Listeria innocua	Nexidia <sup>b</sup>	NEX-1849	Milk	Negative	Not tested
3	Listeria ivanovii	ATCC	BAA-139	Washing water	Negative	Not tested
4	Listeria ivanovii	Nexidia	NEX-923	Duck breast	Negative	Not tested
5	Listeria grayi	CIPc	103321	Food	Negative	Not tested
6	Listeria seeligeri	CIP	100100	Soil	Negative	Not tested
7	Listeria seeligeri	Nexidia	NEX-1848	Milk	Negative	Not tested
8	Listeria welshimerii	CIP	81.49	Decaying vegetation	Negative	Not tested
9	Listeria weihenstephanensis	DSM <sup>d</sup>	24699	Hydrophyte in pond	Negative	Not tested
10	Listeria fleischmanii	DSM	25003	Cheese ripening cellar	Negative	Not tested
11	Lactococcus lactis	ATCC	11454	Reference strain	Negative	Not tested
12	Lactococcus lactis	Nexidia	NEX-363	Food	Negative	Not tested
13	Streptococcus thermophilus	Nexidia	NEX-371	Food isolate	Negative	Not tested
14	Lactobacillus plantarum	ATCC	14917	Pickled cabbage	Negative	Not tested
15	Lactobacillus plantarum	Nexidia	NEX-EC-EnrE8	Surface	Negative	Not tested
16	Lactobacillus buchneri	Nexidia	NEX-EC-E8	Surface	Negative	Not tested
17	Leuconoctoc mesenteroides	NEMIS	KL-LM 18.1	Salad factory	Negative	Not tested
18	Klebsiella pneumoniae	ATCC	4352	Dairy products: cow's milk	Negative	Not tested
19	Aercococcus viridans	NEMIS <sup>e</sup>	MA-LM 191.1	Meat factory	Negative	Not tested
20	Micrococcus yunnanensis	NEMIS	OS-LM 206.2	Meat factory	Negative	Not tested
21	Enterobacter amnigenus	Nexidia	NEX-764	Cocoa	Negative	Not tested
22	Enterobacter cloacae	DSM	30060	Minced meat	Negative	Not tested
23	Cronobacter sakazakii	ATCC	12868	Reference strain	Negative	Not tested
24	Escherichia coli	ATCC	8739	Faeces	Negative	Not tested
25	Citrobacter braakii	ATCC	51113	Snake	Negative	Not tested
26	Pantoea agglomerans	CIP	82.100	Corn crops	Negative	Not tested
27	Salmonella Typhimurium	ATCC	14028	Chicken	Negative	Not tested
28	Pseudomonas aeruginosa	ATCC	15442	Animal room water bottle	Negative	Not tested
29	Providencia vermicola	NEMIS	LU-SM 74E	Egg factory	Negative	Not tested
30	Morganella morganii	NEMIS	LU-SM 74E	Egg factory	Negative	Not tested
31	Enterococcus avium	ATCC	14025	Human feces	Negative	Not tested
32	Enterococcus faecalis	ATCC	51299	Peritoneal fluid	Positive	Negative
33	Enterococcus hirae	ATCC	10541	Reference strain	Positive	Negative
34	Staphylococcus aureus	Nexidia	NEX-1093	Raw milk cheese	Negative	Not tested
35	Staphylococcus aureus	Nexidia	NEX-1069	Mayonnaise	Negative	Not tested
36	Bacillus cereus	CIP	105151	Reference strain	Negative	Not tested
37	Bacillus licheniformis	NEMIS	MH-LM 128.2	Cheese dairy	Negative	Not tested
38	Bacillus subtilis	ATCC	6051	Reference strain	Negative	Not tested
39	Clostridium perfringens	ATCC	13124	Reference strain	Negative	Not tested

<sup>&</sup>lt;sup>a</sup> ATCC: American Type Culture Collection, Manassas, VA, USA. CIP: Collection de l'Institut Pasteur, Paris, 303 France.

<sup>&</sup>lt;sup>b</sup> CIP: Collection de l'Institut Pasteur, Paris, 303 France.

<sup>&</sup>lt;sup>c</sup> CRBIP: Biological Resource Center of institute Pasteur, Paris, France.

<sup>&</sup>lt;sup>d</sup> NCTC: National Collection of 304 Type Cultures, Salisbury, UK.

<sup>&</sup>lt;sup>b</sup> Nexidia: Microbial Strain Collection Nexidia SAS, Dijon, France.

<sup>&</sup>lt;sup>c</sup> CIP: Collection de l'Institut Pasteur, Paris, 303 France.

 $<sup>^{\</sup>rm d}$  DSM: DSMZ-German Collection of Microorganisms and Cell Cultures, Braunschweig, Germany.

<sup>&</sup>lt;sup>e</sup> NEMIS: Microbial Strain Collection NEMIS AG, Au, Switzerland.

Table 3. N-Light *Listeria monocytogenes* vs. Reference (ISO 11290-1:2017) Method – POD Results (4)

		cfuª/		Candidate			Reference			_	
Matrix	Strain	Test Area	N <sup>b</sup>	xc	$POD_C^d$	95% CI	Х	$POD_R^e$	95% CI	$dPOD_{c}^{f}$	95% CI <sup>g</sup>
Stainless Steel	L. monocytogenes ATCCh BAA-751	N/A <sup>i</sup>	5	0	0.00	0.00, 0.43	0	0.00	0.00, 0.43	0.00	-0.43, 0.43
	, 3	70 & 820	20	9	0.450	0.26, 0.66	6	0.30	0.15, 0.52	0.15	-0.14, 0.41
(1" x 1")	& E. faecalis ATCC 29212	450 & 5500	5	5	1.00	0.57, 1.00	5	1.00	0.57,1.00	0.00	-0.43, 0.43

<sup>&</sup>lt;sup>a</sup>cfu/Test area = Results of the cfu/Test area were determined by plating the inoculum for each matrix in triplicate

Table 4. N-Light *Listeria monocytogenes* presumptive vs. confirmed ISO 11290-1:2017 – POD results (4)

	-	cfu <sup>a</sup> /		Candidate presumptive				Candidate o	onfirmed		=
Matrix	Strain	Test Area	$N^b$	Xc	$POD_{CP}^d$	95% CI	Χ	$POD_CC^e$	95% CI	$dPOD_{CP}^{f}$	95% CI <sup>g</sup>
Ctainless Ctaal	L. monocytogenes ATCC <sup>h</sup> BAA- 751 & <i>E. faecalis</i> ATCC 29212	N/A <sup>i</sup>	5	0	0.00	0.00, 0.43	0	0.00	0.00, 0.43	0.00	-0.47, 0.47
Stainless Steel (1" x 1")		70 & 820	20	9	0.45	0.26, 0.66	9	0.45	0.26, 0.66	0.00	-0.13, 0.13
		450 & 5500	5	5	1.00	0.57, 1.00	5	1.00	0.57, 1.00	0.00	-0.47, 0.47

acfu/Test area = Results of the cfu/Test area were determined by plating the inoculum for each matrix in triplicate.

Table 5. N-Light Listeria monocytogenes presumptive vs. confirmed per alternative N-Light Listeria monocytogenes procedure – POD results (4)

	-	cfu <sup>a</sup> /		Candidate presumptive			Candidate confirmed			=	-
Matrix	Strain	Test Area	$N^b$	Xc	$POD_{CP}^d$	95% CI	Х	$POD_{CC}^{e}$	95% CI	$dPOD_{CP}^{f}$	95% CI <sup>g</sup>
Stainless Steel	L. monocytogenes ATCC <sup>h</sup> BAA-751 & E. faecalis ATCC 29212	N/A <sup>i</sup>	5	0	0.00	0.00, 0.43	0	0.00	0.00, 0.43	0.00	-0.47, 0.47
		70 & 820	20	9	0.45	0.26, 0.66	9	0.45	0.26, 0.66	0.00	-0.13, 0.13
(1" x 1")		450 & 5500	5	5	1.00	0.57, 1.00	5	1.00	0.57, 1.00	0.00	-0.47, 0.47

acfu/Test area = Results of the cfu/Test area were determined by plating the inoculum for each matrix in triplicate.

<sup>&</sup>lt;sup>b</sup>N = Number of test portions

cx = Number of positive test portions

<sup>&</sup>lt;sup>d</sup>PODC = Candidate method presumptive positive outcomes confirmed positive divided by the total number of trials

<sup>&</sup>lt;sup>e</sup>PODR = Reference method confirmed positive outcomes divided by the total number of trials

fdPODC= Difference between the confirmed candidate method result and reference method confirmed result POD values

<sup>§95%</sup> CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level

<sup>&</sup>lt;sup>h</sup>ATCC = American Type Culture Collection, Manassas, VA.

 $<sup>^{</sup>i}N/A = Not applicable.$ 

<sup>&</sup>lt;sup>b</sup>N = Number of test portions.

cx = Number of positive test portions.

<sup>&</sup>lt;sup>d</sup>POD<sub>CP</sub> = Candidate method presumptive positive outcomes divided by the total number of trials.

<sup>&</sup>lt;sup>e</sup>POD<sub>CC</sub> = Candidate method confirmed positive outcomes divided by the total number of trials.

fdPOD<sub>CP</sub>= Difference between the confirmed candidate method p0resumptive result and candidate method confirmed result POD values.

<sup>895%</sup> CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level.

<sup>&</sup>lt;sup>h</sup>ATCC = American Type Culture Collection, Manassas, VA.

 $<sup>^{</sup>i}N/A = Not applicable.$ 

<sup>&</sup>lt;sup>b</sup>N = Number of test portions.

cx = Number of positive test portions.

<sup>&</sup>lt;sup>d</sup>POD<sub>CP</sub> = Candidate method confirmed positive outcomes divided by the total number of trials.

<sup>&</sup>lt;sup>e</sup>POD<sub>CC</sub> = Reference method confirmed positive outcomes divided by the total number of trials.

fdPODcp= Difference between the confirmed candidate method result and reference method confirmed result POD values.

<sup>\$95%</sup> CI = If the confidence interval of a dPOD does not contain zero, then the difference is statistically significant at the 5% level.

<sup>&</sup>lt;sup>h</sup>ATCC = American Type Culture Collection, Manassas, VA.

N/A = Not applicable.

#### REFERENCES CITED

- Larose, D., Desroches, N., Reinau, L., Fieseler, L., and Hupfeld, M., Validation of N-Light™ Listeria, L. monocytogenes for the Detection of Listeria monocytogenes
  on environmental surfaces, AOAC Performance Tested Methods™ certification number 122002.
- 2. ISO 11290-1:2017 Microbiology of the food chain -- Horizontal method for the detection and enumeration of *Listeria monocytogenes* and of *Listeria* spp. -- Part 1: Detection method. Geneva, Switzerland
- 3. Desroche, N., De Oliveira, D., Richter, L., Hupfeld, M., Modification for N-Light<sup>™</sup> Listeria monocytogenes Test Kit for Detection of *L. monocytogenes* on Selected Environmental Surfaces, AOAC *Performance Tested Methods*<sup>™</sup> certification number 122002. Approved January 13, 2023.
- Thompson, W., Benzinger, Jr., M.J., Bastin, B., Crowley, E.S., Desroche, N., Ihssen J., and Hupfeld, M., Validation of N-Light™ Listeria monocytogenes for the
  Detection of Listeria monocytogenes on environmental surfaces Level 3 Modification, AOAC Performance Tested Methods<sup>5M</sup> certification number 122002.
  Approved February 9, 2024.